CHANGING THE "NEXT SAFETY CHECK" DATE

Step-by-Step process for VIO 300D ESU

ERBE USA INCORPORATED Surgical Systems

1. Turn the VIO 300D on. Once the *"Safety Check Due"* (green message box) appears, press any tabs ("►" Display Left-side, "◄" Display Right -side) keys to clear the message.

2. Select "Return" option.

3. Select "Other Functions" option.

4. Select "Setup" option.

5. Press the "More" "◄" option until the "Service Program" option is available, select this option.

   The unit will now request a Password. The white buttons (▲= Up, ▼= Down and →= Confirm/Save), that are located to the right of the Front Display, will be used to enter the password.

6. Press the "▼" button once and observe that the "—" (hyphen/dash) now occupies the first character position of the password field.

7. Press the "→" button, this will allow you to gain access to the "Setup" options.

8. Review the current "Date and Time" options; and if required, make the necessary corrections by simply selecting/highlighting the Date and/or Time option, while using the "▼" and "▲" buttons to adjust the settings.

9. Press the More "◄" option several times until the "Next Safety Check" option is available. Select/highlight the option.

10. Press the "▲" button once; confirm that the "Next Safety Check" date adjust to a year from the current date. Press the "→", and turn the unit off.

11. Turn the unit on and confirm that the green message box no longer appears.

    NOTE: The green message will re-appear a year from this update.